



## URGENT FIELD SAFETY NOTICE

GE Healthcare

3000 N. Grandview Blvd. - W440  
Waukesha, WI 53188, USA

<Date of Letter Deployment>

GEHC Ref# 76177-2

To: Hospital Administrators /Risk Manager  
Biomedical Engineering  
Head of Primary Care Ultrasound Department

RE: Use of Tristel Trio Wipes System with GE TEE probes

**Please ensure that all potential users in your facility are made aware of this safety notification and the recommended actions.**

If your practice is to follow the GE TEE care card for the disinfection of your TEE probes, please continue this practice and this letter does not apply. This letter addresses customers who are using the Tristel Trio Wipe System as their method of disinfecting their TEE probes.

The Tristel Trio Wipes System has been validated for efficacy and residual toxicity for TEE probes that do not have visible signs of wear or damage, and when used with shorter procedure times such as cardiac scans. Note that the Tristel Trio Wipe System is not approved or available in all countries or regions.



**Safety Issue**





Further testing has identified a need for clarification of previously communicated disinfection guidance for GE TEE Probes.

Specifically, in-use experience and limited testing has revealed that the previously communicated disinfection methods using the Tristel Trio Wipe System as a method of disinfecting TEE probes, may not always be effective when used in connection with procedures that have long TEE probe intubation times, such as cardiac surgeries, and/or with TEE probes with visible signs of wear or damage in areas to be disinfected, both of which may increase the bioburden on the probe. Therefore, under these conditions, GE recommends the use of a sterile sheath covering the TEE probe in addition to disinfection using the Tristel Trio Wipes System as described below.

**Safety Instructions**

As part of each use of the probe when using the Tristel Trio Wipes System is appropriate (as per the guidance above), do the following:

	
<p><b>STEP 1 of 4:</b> Process the TEE probe immediately after extraction from the patient. Do not allow bodily fluids to dry on the probe.</p>	<p><b>Step 2 of 4: Pre-Clean Wipe</b> Thoroughly wipe the surface until soil and organic matter have been visibly removed. Use at least two wipes.</p> <p><i>Note: These instructions may differ from the manufacturer's guidance provided with the wipes, in that two pre-cleaning wipes must be used every time rather than one. This is to ensure</i></p>

			<i>an effective enzymatic cleaning, prior to disinfection.</i>
			
<b>Step 3 of 4: Sporocidal Wipe</b> Apply Tristel activator foam according to the instructions on the bottle and in the accompanying instruction leaflet from Tristel. Scrunch the wipe for 15 seconds. Ensure it is evenly covered with foam. Wipe the endoscope until it has been covered with Tristel. Wait for at least 30 seconds.			<b>Step 4 of 4: Rinse Wipe</b> Thoroughly wipe the surface that has been decontaminated.

**Affected Product Details**

The following TEE probe models: 6VT-D, 6Tc, 6Tc-RS, 6T, 6T-RS, 9T and 9T-RS.

*Table 1: UDI marked probes shipped after July 2016*

Probe type	Part number	GTIN
6VT-D	KN100120	00840682115650
6VT-D	KN100110	00840682115681
6Tc	KN100107	00840682115582
6Tc-RS	KN100106	00840682115735
9T	KN100121	00840682115636
9T-RS	KN100122	00840682115728

*Table 2: Non-UDI marked probes shipped prior to July 2016*

Probe type	Part number
6VT-D	KN100120
6VT-D	KN100110
6VT-D	KN100100
6Tc	KN100107
6Tc	KN100105
6Tc-RS	KN100104
6Tc-RS	KN100106
9T	KN100121
9T	KN100072
9T-RS	KN100122
9T-RS	KN100073
6T	KN100092
6T	KN100094
6T	KN100068
6T	KN100022
6T-RS	KN100093
6T-RS	KN100095
6T-RS	KN100062

The probe part numbers and types are printed on the rating plate of the probe connector.

**Contact  
Information**

If you have any questions or concerns regarding this notification, please contact GE Healthcare Service or your local Service Representative.

Please be assured that maintaining a high level of safety and quality is our highest priority. If you have any questions, please contact us immediately per the contact information above.

Sincerely,



James W. Dennison  
Vice President - Quality & Regulatory  
GE Healthcare



Jeff Hersh, PhD MD  
Chief Medical Officer – Medical Safety  
GE Healthcare