

URGENT MEDICAL DEVICE FIELD SAFETY NOTICE

GRIPPER® and PORT-A-CATH® Needles Shipped Without Current EC Certification

Affected Device: GRIPPER® and PORT-A-CATH® Needles

Type of Action: Removal

Date: August 10, 2018

Attention: Clinical Users of, and Distributors of affected GRIPPER® and PORT-A-CATH® Needles

Affected Devices: The following Product Numbers and associated Lots affected by this issue are listed in Table 1 below:

Table 1

Product Number	Lot Number	Product Number	Lot Number
21-2014-24	3550725, 3587313	21-2861-24	38X082
21-2714-24	3595538	21-2865-24	38X130
21-2715-24	38X153	21-2866-24	38X090
21-2718-24	38X254	21-2868-24	38X010
21-2720-24	3603299	21-2948-24	38X248
21-2733-24	3599339	21-2949-24	38X055
21-2734-24	3599337, 3599337	21-3271-24	38X059
21-2736-24	3595533	21-3277-24	38X004
21-2767-24	3618227	21-3362-24	38X261
21-3367-24	3599273	21-3363-24	38X278

Dear Customer,

The purpose of this Field Safety Notice (FSN) is to advise you that Smiths Medical has initiated a voluntary Field Safety Corrective Action (FSCA) for specific GRIPPER® and PORT-A-CATH® Needles. A total of 29,292 devices are included in this FSCA. Product and lot number information of affected product in your possession can be found on the Response Form accompanying this FSN.

REASON FOR FIELD SAFETY CORRECTIVE ACTION:

Smiths Medical became aware that the devices listed in Table 1 were inadvertently shipped to European countries without current EC Certification.

Representative images of affected device and examples of affected device labeling are shown in Figure 1.

