

## Urgent FIELD SAFETY NOTICE

Device: **Terumo® Hypodermic NEOLUS Needle – Incidental Compromised Package Integrity**

Reference: **FSN 1905 2019-09**

Action: **Return**

---

**Attention:** Chief of Hospital, Clinics, Pharmacy & Medical staff

### Description of the problem

Terumo Europe has become aware through its internal investigation of a risk for incidental occurrences of compromised packaging integrity, potentially compromising the product sterility, in a defined population of Terumo® Hypodermic NEOLUS Needle.

The conclusion of the current investigation indicates that in a specified period, a temporary technical issue on the packaging line caused pressure on the blister film. This resulted in occasional damages of the film, potentially impacting the integrity of the blister in the affected population.

Although a portion of the affected population was previously released to the market being conform to specifications, the investigation could not fully rule out potential occurrences of individual affected packages in this population.

Terumo Europe is voluntarily conducting this Field Safety Corrective Action for the affected device population as a precautionary measure.

There is no reported complaint from the market regarding this issue.

### Details on affected devices

| Product code | Description                      | Affected lot numbers |
|--------------|----------------------------------|----------------------|
| NN-2525R     | Terumo® Hypodermic NEOLUS Needle | 1907005              |
| NN-2525R     | Terumo® Hypodermic NEOLUS Needle | 1907006              |
| NN-2316R     | Terumo® Hypodermic NEOLUS Needle | 1907007              |
| NN-2525R     | Terumo® Hypodermic NEOLUS Needle | 1907009              |
| NN-2316R     | Terumo® Hypodermic NEOLUS Needle | 1908002              |
| NN-2525R     | Terumo® Hypodermic NEOLUS Needle | 1906012              |
| NN-2525R     | Terumo® Hypodermic NEOLUS Needle | 1906013              |

### Potential hazard

The possibility of a compromised device being used is considered low. Accidental use of a compromised device may occasionally result in health consequences for the patient, such as infection.

### Corrective actions

Terumo Europe has been implementing corrective actions assuring elimination of such remote defect.

Terumo Europe is alerting its involved customers about the issue, and is asking to immediately identify, segregate and return the remaining units in their inventory to Terumo Europe.

### Customer instructions

- 1) Review this Field Safety Notice and assure that all users are aware of this notice.
- 2) Immediately identify and segregate the units of the suspected device population.
- 3) Indicate the number of unused units from the referred codes/lots on the related reply form and return this form as quickly as possible to the e-mail address or the fax number indicated on the form.
- 4) The company representative will contact you to organize immediate pick-up and provide replacement.

We confirm that this *Field Safety Notice* has also been notified to your national Competent Authority.

We encourage you to contact us or your local Terumo representative with any questions or concerns.

Organisation (to be completed by the sales or dealer)  
Contact name (function)  
Contact phone, mobile, email



Fayeze Abou Hamad  
MD Vigilance Expert  
Terumo Europe NV –  
Leuven, Belgium

