
Supplement to Urgent safety information
Recall
concerning
Mediware Infusion Set, Ref.- no. H7 0303

06.04.2021

Sender:

Servoprax GmbH
Am Marienbusch 9
46485 Wesel

Adressed to:

Users, Distributors, Risk Managers, Medical device safety officers

Supplement: Another batch of article H7 0303 is also affected by this recall:

REF. H7 0303 Batch: 17A050 /expiry date: 31.12.2021.

Please check whether you still have stocks of this batch and stop using these products as a precaution.

Identification of the Medical Devices concerned:

| Article Number (Ref.-no.) | Article description |
|--|---|
| H7 0303 | Mediware Infusionengerät steril PERFUD.CP-FLEX , mit Belüftung Mediware Infusion Set, sterile |
| The following batches are affected: | |
| Batch | Expiry date |
| 16F080 | 31.05.2021 |
| 16F081 | 31.05.2021 |
| 17A050 | 31.12.2021 |
| 17A051 | 31.12.2021 |
| 17K028 | 31.10.2022 |
| 19A065 | 31.12.2023 |

As a precaution, please stop using the above products with IMMEDIATE EFFECT.

Description of the problem including the identified cause:

Dear Sir or Madam ,

We would like to inform you with this letter about a precautionary product recall of Mediware Infusion Set, Ref.- no. H7 0303.

We have been informed that significant quality problems have occurred in the sterilization company where the infusion devices were sterilized.

At this point in time, we do not know whether our articles are actually affected by this incident. We have not received any complaints or feedback from our customers regarding this product.

However, to ensure patient safety, we are recalling the above batches as a precaution.

The batches that we currently have in stock were sterilized in another specialised company, do not have any defects and are not affected by this process.

Servoprax GmbH
Am Marienbusch 9
46485 Wesel

What actions are to be taken by the addressee?

Please carry out the following actions:

- Ensure that this safety information is read and understood by the persons in your organisation who use the above mentioned Infusion Sets.
- If you have supplied the product to third parties, please forward a copy of this information to the customers or users concerned.
- Identify the affected batches of the product in your institution/company and ensure that a stock of the products is completely withdrawn from any kind of use. Please destroy any existing batches.

Please complete the attached customer response form and return it to servoprax GmbH as soon as possible.

If you no longer have any stock of the above products, please also indicate this on the reply form.

Contact person:

If you have any questions or uncertainties in connection with this Safety Notice, please contact the Quality Management Department at the following address

Ms Maria Mohrmann (Safety Officer for Medical Devices)

Mail: maria.mohrmann@servoprax.de / Tel. 0281 95283-51.

or to our complaints department, Ms Lisa Spodick

Mail: lisa.spodick@servoprax.de / Tel. 0281 95283-25

Thank you for your attention and assistance.

Yours sincerely

servoprax GmbH
Maria Mohrmann