

## **Urgent FIELD SAFETY NOTICE**

Device: **Terumo® Extracorporeal blood circuits, Tubing sets and Ancillary devices**

Reference: **FSN 2102 -- 2021-05**

Action: **Inspect or Return**

**Attention:** Chief of Perfusion; Director of Operating Room Services; Director of Biomedical Services; Risk Management

### **Description of the problem**

Terumo Europe has received two complaints for Extracorporeal blood circuits, in which, whilst opening the packaging of the device, the perfusionist detected the outer bag was not sealed.

Further investigation into the seals determined that on a specific sealer for a defined period of time, a full weld had not been obtained between the two films, resulting in a number of units with a seal which can be peeled open if during handling the weight of the product is transferred onto the sealed end of the packaging.

There was no impact on patients as this failure was detected before use, however, Terumo Europe is conducting this voluntary Field Safety Corrective Action (FSCA) as a precautionary measure.

### **Details on affected devices**

Product code	Product Description	Lot Numbers
<b>CX-BE068X</b>	NEONATE PACK WITH FX05RW IN YELLOW BIN	1905120, 1908484
<b>CX-BE093X</b>	TUBING SET WITH FX15RW40C IN YEL BIN	2005151, 2005317
<b>CX-BE094X</b>	ADULT PACK WITH FX25RWC IN YEL BIN	1903203, 1904220, 1905362, 1907097
<b>CX-CH088X</b>	ADULT TUBING PACK WITH FX25E & R400C	1903126, 1903442, 1904341, 2004273
<b>CX-CH094X</b>	ADULT PACK WITH PRECON FX15RE30C	1903224, 1904103, 1905398
<b>CX-CH132X</b>	ADULT PACK XCOAT WITH PRECONNECT FX25REC	1904100, 1905104, 1906111, 1908344, 2004157, 2005231, 2005305
<b>CX-CZ095XA</b>	ADULT SET FX25RWC PACKED IN YELLOW BIN	1904375, 1904377, 1908370, 1908371, 2004022, 2004197, 2004198, 2004211, 2004212
<b>CX-CZ095X</b>	ADULT SET W/ FX25RW PACKED IN YELLOW BIN	1904297
<b>CX-CZ096X</b>	ADULT SET W/ FX25RW PACKED IN YELLOW BIN	1908495
<b>CX-DK098</b>	SMALL ADULT PACK PRECON FX15RW30 YEL BIN	1903171, 1905176, 2005216
<b>CX-GE279</b>	TUBING SET PRECONNECTED FX25REC	1904135, 1905122, 1905211, 2003296
<b>CX-UK484</b>	LARGE CHILD SET WITH FX15RW30 YEL BIN	1905295

<b>CX-UK575</b>	ADULT PACK WITH FX25REC	1903187, 1903271, 1904262, 1905212, 1907040, 1908451, 2005330
-----------------	-------------------------	---

### Potential hazard

All observations from complaints and testing indicate a clearly identifiable breach of the seal.

It is highly unlikely than an open pack remains undetected during the un-packing process at the hospital; therefore, an open pack is unlikely to be used thereby excluding adverse health consequences to patients.

### Corrective actions

The specific sealer was found being used for a defined period of time outside of its validated temperature range and is removed out of service.

### Customer instructions

1. Review this communication and ensure that all users have received notice of this issue.
2. This notice needs to be passed on to any organization where the potentially affected products have been transferred.
3. Complete the related reply form and return this form as quickly as possible to the e-mail address indicated on the form.
4. In case, because of medical need, the option for inspection of the devices is chosen, follow the instructions in Annex.
5. Your local Terumo representative will contact you and further follow up depending on your choice for device inspection or device return.

We confirm that this *Field Safety Notice* has also been notified to your national Competent Authority.

We encourage you to contact us or your local Terumo representative with any questions or concerns.

Organisation (to be completed by the sales or dealer)  
Contact name (function)  
Contact phone, mobile, email



Fayez Abou Hamad  
MD Vigilance Expert  
Terumo Europe NV –  
Leuven, Belgium

## Annex – instructions for seal inspection

The following instructions below are to be used to remove the tubing sets from within the yellow bins.

1. On receipt of the packs please carefully open the outer packaging.
2. Remove the cardboard interleave and any documentation.
3. Once the cardboard interleave has been removed there will be a tubing set sealed within a gas bag placed within a yellow bin as shown in image 1 below.



Image 1 – Pack when opened

4. Remove the pack carefully by grabbing the plastic insert within the Primary packaging. Place one hand either side and pull by gripping both the product (within the plastic insert) and the packaging together. It is critical that you do not remove the pack by pulling only on the packaging. Image 2 shows how the pack is to be removed.



Remove the pack by placing one hand either side and pull by gripping both the product (within the plastic insert) and the packaging together as shown in image 2.

Image 2 – How to grip packs when removing

5. Once the packs have been removed place on their side and inspect packaging. The seal on the bottom of the pack should be intact as shown in image 3 below. If the pack has been removed by pulling on the gas bag without holding the plastic insert it is possible that the seal has come open (as shown in image 4). In this case the failure of the seal will be a wide opening of 10 – 15 cm.



Image 3 – Pack as seen when removed from yellow bin

Inspection – Inspect the bottom seal on removal: it should be sealed.

In the case that the seal is open the pack is to be discarded. An example of an open seal is shown in image 4 below.



Image 4 – Example of an open seal.

## Field Safety Notice - CUSTOMER REPLY FORM

Device: **Terumo® Extracorporeal blood circuits, Tubing sets and Ancillary devices**

Reference: **FSN 2102 -- 2021-05**

Action: **Inspect or Return**

Please complete, sign and e-mail this form back:

To:

E-mail:

<b>Hospital/Customer Name</b>		
<b>City</b>		
<b>Country</b>		
<p><b>Our records indicate that you have received devices from the involved device population.</b></p>		
<p><b>By completion and return of this form, I am confirming receipt, reading and acting on this Field Safety Notice:</b></p> <p><input type="checkbox"/> We have no physical inventory of the involved devices</p> <p><input type="checkbox"/> We will inspect the products and further use them in case not affected</p> <p><input type="checkbox"/> We will return the devices and the following devices are ready to return</p>		
<b>Reference</b>	<b>Lot number</b>	<b>Number of units ready to return</b>
<b>Person Responding [Please Print]</b>		
<b>Title</b>		
<b>Phone Number</b>		
<b>Signature</b>		
<b>Date</b>		

FSN2102A [EN]