Helena Biosciences Notification Urgent Field Safety Notice



Electrophoresis Division

CEP028 25 FEB 2022

Dear Colleague,

Further to the previous technical alert CEP026 issued 24-DEC-2021 regarding sub-component REF: 0012.075 Bottom sheet rack/Rack Loading Plate (in relation to the V8 Nexus REF: 800008, 800018, and 800020) an investigation, and a revised risk assessment has been conducted internally.

Based on the outcome of these, further corrective actions are not required as the risk assessment has identified the probability of harm as low.

However, users are advised to be cautious when handling and cleaning the instrument and as per standard good laboratory practice, please ensure appropriate PPE is worn at all times when operating/handling/cleaning the instrument. Please refer to the current V8 Operator Manual (HL-2-P-2219 Rev 5).

Email:

Technical Support v8support@helena-biosciences.com

Tel: +44 (0)191 4828440