

## Medtronic (Schweiz) AG

Talstrasse 9  
3053 Münchenbuchsee  
www.medtronic.com

Tel. 031 868 01 00  
Fax 031 868 01 99  
E-Mail [swisscontact@medtronic.com](mailto:swisscontact@medtronic.com)

Helpline:: 0800 633 333

## Field Safety Notice MiniMed™ 600 and 700 series insulin pump Battery Cap Notification

Insulin Pump	Model Number
MiniMed™ 640G Insulin Pump	MMT-1711, MMT-1712, MMT-1751, MMT-1752
MiniMed™ 670G Insulin Pump	MMT-1761, MMT-1762, MMT-1781, MMT-1782
MiniMed™ 720G Insulin Pump	MMT-1809, MMT-1810, MMT-1859, MMT-1860
MiniMed™ 740G Insulin Pump	MMT-1811, MMT-1812, MMT-1861, MMT-1862
MiniMed™ 770G Insulin Pump	MMT-1881, MMT-1882, MMT-1891, MMT-1892
MiniMed™ 780G Insulin Pump	MMT, 1885, MMT-1886, MMT-1895, MMT-1896

May 2022

Medtronic Reference: FA1249

Dear Pump User,

You are receiving this letter because our records indicate that you have a MiniMed™ 600 series and/or MiniMed™ 700 series insulin pump. For your safety, we want to inform you of a potential issue relating to your pump's battery cap and provide actions you should take. Please carefully review the information below.

### **ISSUE DESCRIPTION**


The battery cap on the pump consists of a plastic cap and a metal contact that work together with the AA battery to power the pump. The metal contact should be held in place by three raised, round, black, plastic dots, as pictured below. If the metal contact becomes loose or falls off from the battery cap, it can result in an incomplete battery connection, leading to no power source to the pump. When the pump detects no power source, an "Insert battery" alarm will occur, and **insulin delivery will immediately stop**. After 10 minutes, the alarm sound may increase to a siren, and **the pump will turn off**.



The battery cap can be found on the top of the pump where the AA battery compartment is located



Undamaged battery cap - Continue to use  
Has three raised, round, black, plastic dots holding metal contact in place



Sample screen image of "Insert battery" alarm



Damaged battery caps - Do not use  
Metal contact is missing, or fewer than 3 raised dots can be seen

If the pump stops delivery of insulin due to power loss, this could lead to varying degrees of high blood sugar, including Diabetic Ketoacidosis (DKA). Serious injuries have been reported with the use of the MiniMed™ 600 series and MiniMed™ 700 series insulin pumps associated with the damaged cap, but not all have been directly correlated to this issue based on review with independent clinical experts. Damaged battery cap contacts could potentially lead to those events as explained above.

### **ACTIONS REQUIRED**

*Before you begin: Do not remove the battery cap unless you have a new battery available. If you have a spare undamaged battery cap, ensure it is available nearby.*

**During routine battery replacement, check the metal contact on your pump battery cap** to see if it is loose, damaged, or missing. Do not try to lift or move the metal contact upon inspection (see picture above).

- **If the battery cap contact is not damaged**, continue to use your pump and monitor for cap damage during battery replacement.
- **If the battery cap contact is damaged**, immediately replace it with a spare cap that you may have received with your original pump shipment, and discard the damaged cap. If you do not have a spare cap, stop using your pump and revert to a back-up plan per your healthcare provider's recommendations. Then, contact our Helpline / your Medtronic contact at 0800 633 333 to request a spare battery cap.
- **If you are unsure if the battery cap contact is damaged**, replace it with a spare cap or contact our Helpline / your Medtronic contact at 0800 633 333.
- Always pay close attention to the pump and pump battery status after inserting the new battery.

**Medtronic Actions:**

We are working on a new design for the cap and we will keep you updated when we ship one once it is approved and available for use. We are committed to continuously monitoring and improving your experience with our products and will proactively share important safety updates.

We understand this impacts your experience and are here to support you. If you have further questions, please call our Helpline / your Medtronic contact at 0800 633 333.

Sincerely,

**Medtronic (Schweiz) AG**