

Atellica® Solution
ADVIA Centaur® XP
ADVIA Centaur® XPT
ADVIA Centaur® CP

Atellica IM and ADVIA Centaur Systems Urine Cortisol Negative Bias

Our records indicate that your facility may have received the following product:

Table 1. Atellica® Solution and ADVIA Centaur® Systems Affected Product(s)

Assay	Siemens Material Number (SMN)	Unique Device Identification (UDI)	Kit Lot #	Exp. Date (YYYY-MM-DD)	Mfg. Date (YYYY-MM-DD)
Atellica IM Cortisol 50T	10995538	(01)00630414598659(10)50732343(17)20230707 (01)00630414598659(10)50733343(17)20230707	50732343 50733343	2023/07/07	2022/04/07
		(01)00630414598659(10)88975345(17)20230930	88975345	2023/09/30	2022/06/30
		(01)00630414598659(10)29106347(17)20240103	29106347	2024/01/03	2022/10/03
		(01)00630414598659(10)29107347(17)20240103	29107347 and higher	2024/01/03	2022/10/03
Atellica IM Cortisol 250T	10995537	(01)00630414598642(10)50734343(17)20230707	50734343	2023/07/07	2022/04/07
		(01)00630414598642(10)88974345(17)20230930	88974345	2023/09/30	2022/06/30
		(01)00630414598642(10)29105347(17)20240103	29105347 and higher	2024/01/03	2022/10/03
ADVIA Centaur Cortisol 50T	10994924	(01)00630414602943(10)49741344(17)20230707	49741344	2023/07/07	2022/04/07
		(01)00630414602943(10)07241346(17)20230930	07241346	2023/09/30	2022/06/30
		(01)00630414602943(10)88318346(17)20230930	88318346	2023/09/30	2022/06/30
		(01)00630414602943(10)29109348(17)20240103	29109348 and higher	2024/01/03	2022/10/03
ADVIA Centaur Cortisol 250T	10994926	(01)00630414602950(10)49742344(17)20230930	49742344	2023/07/07	2022/04/07
		(01)00630414602950(10)88319346(17)20230930	88319346	2023/09/30	2022/06/30
		(01)00630414602950(10)29108348(17)20240103	29108348 and higher	2024/01/03	2022/10/03

This issue affects all current and future lots of the Atellica IM and ADVIA Centaur Systems Cortisol assays until the product issue is resolved.

Reason for Correction

The purpose of this communication is to inform you of an issue with the product indicated in Table 1 above and provide instructions on actions that your laboratory must take.

Siemens Healthcare Diagnostics Inc. has confirmed customer observations of a negative bias with urine patient samples and urine cortisol Quality Control (QC) results intermittently out of range low. As indicated in the Instructions for Use, if QC results do not fall within with the assigned values, do not report results.

Lot-to-lot comparison studies between affected and unaffected kit lots showed a negative bias with urine samples (see Figure 1). Siemens is currently investigating the root cause of the negative bias and working to restore assay performance with urine samples.

Siemens requires that customers discontinue use of urine patient samples with kit lots listed in Table 1 and higher until further notice.

Serum and plasma samples are **NOT** affected, and customers can continue to use the Cortisol assay with these sample types.

Risk to Health

Worst case, there is a potential for erroneously depressed urine Cortisol patient results, which may lead to a delayed differential diagnosis of hypercortisolism, such as Cushing Syndrome. Mitigations include correlation of results to patient's clinical signs and symptoms, additional laboratory and repeat testing findings.

Actions to be Taken by the Customer

- Discontinue use of urine patient samples with the Atellica IM Cortisol and ADVIA Centaur Cortisol assays until further notice.
- Customers may continue to use serum and plasma sample types with the Atellica IM Cortisol and ADVIA Centaur Cortisol assays.
- Please review this letter with your Medical Director to determine the appropriate course of action, including for any previously generated results, if applicable.
- Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
- If you have received any complaints of illness or adverse events associated with the products listed in Table 1, immediately contact your local Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.


Please retain this letter with your laboratory records and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

Additional Information

One hundred urine samples were tested with Atellica IM Cortisol kit lots ending in 341 (unaffected) and 345 (affected) across the Assay Measuring Range. Figure 1 shows the observed % differences (bias). This data is representative of all in-date Atellica IM lots. While the investigation showed the ADVIA Centaur XP/XPT and ADVIA Centaur CP have a lesser degree % difference, these platforms are included in this Field Action because they are manufactured with the same materials.

Figure 1. % Difference Plot – Atellica IM Cortisol Kit Lots Ending in 345 vs Kit Lots Ending in 341



Atellica Solution and ADVIA Centaur Systems are trademarks of Siemens Healthcare Diagnostics Inc.

Legal Manufacturer SRN: US-MF-000016560

FIELD CORRECTION EFFECTIVENESS CHECK

Atellica IM and ADVIA Centaur Systems Urine Cortisol Negative Bias

This response form is to confirm receipt of the enclosed Siemens Healthcare Diagnostics Urgent Field Safety Notice AIMC 23-07.A.OUS dated March 2023 regarding Atellica IM and ADVIA Centaur Systems Urine Cortisol Negative Bias. Please read the question and indicate the appropriate answer.

Return this completed form to Siemens Healthcare Diagnostics as per the instructions provided at the bottom of this page.

1. I have read and understood the Urgent Field Safety Notice instructions provided in this letter. Yes ☐ No ☐

Name of Person Completing Questionnaire: _____

Title: _____

Institution: _____

Instrument Serial Number: _____

Street: _____

City: _____

State: _____

Phone: _____

Country: _____

Please send a scanned copy of the completed form via email to XXXX@XXXX.

Or to fax this completed form to the Customer Care Center at XXXXXX.

If you have any questions, contact your local Siemens Healthineers technical support representative.