

Field Safety Notice

Date: 30 November 2023

Field Safety Notice Number: RFA-2A

Dear Customer,

Our records indicate that we have supplied your organisation with **The Insides Driver (Driver)**. The Driver is a component of The Insides System and The Insides Patient Education Model.

As you may be aware, on the 12th of October a Driver overheated at a hospital in Europe. We traced it back to a small number of units that contained a faulty internal battery from a 3rd party supplier. We then voluntarily and urgently recalled all of these affected units from the field, while continuing with a full investigation.



The Insides Driver

We have now concluded this investigation. This has identified opportunities to add additional safety mechanisms to the Driver, and these additions are currently being implemented in a new version of the device.

As a result of this investigation and with an abundance of caution, we have decided to voluntarily recall all remaining Drivers, and replace them with a new version of the Driver that contains the additional safety mechanisms.

We do not perceive there to be any risk in patients continuing to perform chyme reinfusion with their existing Drivers (with the exception of the initial urgently recalled devices). Please note that the immediate cessation of patient use of the device without replacement could put patients at risk of dehydration, hospitalisation, and invasive procedures (eg: central line placement).

The Insides Company will supply **Driver Replacement Kits** to make this replacement as easy as possible. Each kit contains a new Driver, an updated IFU, and replacement instructions.

The Company is doing this to ensure all patients, healthcare professionals, and our distribution partners have access to this new version of the Driver with the additional safety mechanisms. We expect the Driver Replacement Kits will be available and ready to ship in December.

Field Safety Notice

Date: 30 November 2023

Field Safety Notice Number: RFA-2A

Our Requested Actions for all Units of The Insides Driver listed on Page 6 are

1. Please communicate this Field Safety Notice to all relevant staff and users.
2. Where applicable, please communicate this Field Safety Notice to third parties you have supplied Drivers to.
3. Please complete the **Customer Response Form**, provided as an editable attachment, and email it to:

Tina Mason: tina.mason@theinsides.co
Garth Sutherland garth.sutherland@theinsides.co
4. Upon receipt of the completed Customer Response Form, The Insides Company will arrange Driver Replacement Kits.
5. Once you receive the Driver Replacement Kits, please replace the following Drivers:
 - a. Unused Drivers contained in The Insides System, and
 - b. Drivers that are in use with a first patient (as Drivers are Single Patient Use Only).
6. At this time please also collect all used Drivers and Patient Education Models (PEMs).
7. Return all used Drivers, Drivers that have been replaced and Patient Education Models (PEMs) to the return address specified on the Customer Response Form.

We are seeking assistance from our Distribution partners and direct customers to replace all Drivers at their premises, and at their hospitals.

Patients using their Drivers at home can have them replaced by hospital staff when the patient returns to hospital for their monthly tube change.

Drivers that have been used to complete treatment on a single patient will not be replaced, as all Drivers are Single Patient Use only, and must not be used on more than one patient.

Field Safety Notice

Date: 30 November 2023

Field Safety Notice Number: RFA-2A

Given that the Driver is designated for single patient use, it's probable that the majority of devices have been made inactive and discarded following the completion of chyme reinfusion therapy by the patients.

Patient Education Models (PEMs) also contain Drivers. These will also need to be returned to have their Drivers replaced.

The Insides Company has temporarily paused supplying any new consignments of The Insides System, until the new Drivers are available. As soon as the new Driver is available all outstanding orders will be fulfilled with priority.

We sincerely regret the inconvenience that this Field Safety Notice will cause. We greatly appreciate your understanding as we take actions to prioritise, and healthcare professional safety.

The undersigned confirms that the Field Safety Notice has been notified to the appropriate Regulatory Agency. Should you have questions or need additional information, please contact The Insides Company.

Kindest Regards



Tina Mason
Quality, Regulatory & Sustainability Manager
The Insides Company
tina.mason@theinsides.co



Garth Sutherland
CEO
The Insides Company
garth.sutherland@theinsides.co

Field Safety Notice

Date: 30 November 2023

Field Safety Notice Number: RFA-2A

Existing Driver and Replacement Driver (v4)



Existing Drivers
have a
WHITE Label



New (Replacement) Drivers
have a
PURPLE Label

Driver Replacement Kits and Contents




The Insides Driver v4



Returned Device Sticker



The Insides System IFU v 25.0






Replacement Instructions

Field Safety Notice

Date: 30 November 2023

Field Safety Notice Number: RFA-2A

Manufacturer Single Registration Number (SRN)	NZ-MF-000011306
Product Affected	<p>The Insides Driver Part Number: PS001(Rest of World) & PS009 (US)</p>  <p>The Insides Driver is a component of:</p> <p>The Insides System Part Numbers: PS004 = US; PS005 = UK/EU and PS006 = NZ/AU/ZA)</p>  <p>The Insides System Patient Education Model (PEM) Part Number: MK-1R001</p> 
Product Issue	<p>As you may be aware, on the 12th of October a Driver overheated at a hospital in Europe. We traced it back to a small number of units that contained a faulty internal battery from a 3rd party supplier. We then voluntarily and urgently recalled all of these affected units from the field, while continuing with a full investigation.</p> <p>We have now concluded this investigation. This has identified opportunities to add additional safety mechanisms to the Driver, and these additions are currently being implemented in a new version of the device.</p>
Resolution	<p>As a result of this investigation and with an abundance of caution, we have decided to voluntarily recall all remaining Drivers, and replace them with a new version of the Driver that contains the additional safety mechanisms.</p> <p>The Company is doing this to ensure all patients, healthcare professionals, and our distribution partners have access to this new version of the Driver with the additional safety mechanisms.</p>

Field Safety Notice

Date: 30 November 2023

Field Safety Notice Number: RFA-2A

Affected Units	The Insides System	The Insides Driver*
	PS004 The Insides System US LOT Range: LOT0000150 - LOT0000277 SN Range: 04000001 - 0400002D UDI: 09421905447065	PS001 The Insides Driver (ROW) LOT Range: LOT0000036 - LOT0000281 SN Range: 01000001 - 0100032C UDI: 09421905447010
	PS005 The Insides System (UK/EU) LOT Range: LOT0000038 - LOT0000274 SN Range: 05000001 - 0500016D UDI: 09421905447034	PS009 The Insides Driver US LOT Range: LOT0000134-LOT0000276 SN Range: 09000001 - 09000023 UDI: 09421905447089
	PS006 The Insides System (NZ/AU/ZA) LOT Range: LOT0000039 -LOT0000275 SN Range: 06000001 - 060000AB UDI: 09421905447058	

* Individual Driver LOT & SN details are provided in addition to the System details as if these are in use in the field, access to the corresponding System LOT & Serial numbers may no longer be available in the event the System packaging was disposed of.

Drivers that are part of a Patient Education Model (PEM) will not have a System associated with them.

Customer Response Form

Reference: Field Safety Notice: RFA-2A

Our records indicate that the The Insides Driver (included in PS005/PS006 The Insides System and Patient Education Model) was provided to you.

Important: The Customer Response Form provides The Insides Company with the means to monitor the progress of Field Actions. Please complete this form regardless of whether the affected product is still at your facility. It is imperative that you return this form for our records.

Field Safety Notice Receipt and Customer Response Form Completion	
Customer action undertaken on behalf of Distributors	
<input type="checkbox"/> I acknowledge receipt of the Field Safety Notice dated 13 November 2023 related to the above product. We have read and we understand the communication and the required actions.	Distributor complete or enter N/A
<input type="checkbox"/> I have checked my stock	Distributor to enter quantity
<input type="checkbox"/> I have identified customers that received or may have received this device	
<input type="checkbox"/> I have attached customer list	
<input type="checkbox"/> I have informed the identified customers of this FSN	Date of communication:
<input type="checkbox"/> I have received confirmation of reply from all identified customers	
<input type="checkbox"/> Neither I nor any of my customers has any affected devices in inventory	

I confirm I have the following devices and requires a replacement as specified in the Field Safety Notice.
Please add more rows as required.

PRODUCT PS001 or PS005 or PS006	Serial Number	CURRENT LOCATION
TOTAL AFFECTED PRODUCT		

Customer Response Form

Reference: Field Safety Notice: RFA-2A

Organisation			
Contact Name / Title			
Address (no PO boxes, please)			
City, State, Zip			
Phone Number		Fax:	
E-Mail Address:			
Signature		Date	

Address to Return Product To:

JAN KREDIET VENLO
Jacob Roggeveenweg 8-10
5928 LS Venlo
Netherlands

For the attention of: The Insides Company

You can request a courier bag from www.theinsides.co/returns page and we will send you a courier bag/box for returning the old unit. We will responsibly dispose of these units.

Please return your completed form by email to:

tina.mason@theinsides.co and garth.sutherland@theinsides.co

Customer Response Form

Reference: Field Safety Notice: RFA-2A

Our records indicate that the The Insides Driver (included in PS005/PS006 The Insides System and Patient Education Model) was provided to you.

Important: The Customer Response Form provides The Insides Company with the means to monitor the progress of Field Actions. Please complete this form regardless of whether the affected product is still in use at your facility. It is imperative that you return this form for our records.

Field Safety Notice Receipt and Customer Response Form Completion	
Customer action undertaken on behalf of Healthcare Organisation	
<input type="checkbox"/> I acknowledge receipt of the Field Safety Notice dated 13 November 2023 related to the above product. We have read and we understand the communication and the required actions.	Customer complete or enter N/A
<input type="checkbox"/> I performed all actions requested by the Field Safety Notice.	Customer complete or enter N/A
<input type="checkbox"/> The information and required actions have been brought to the attention of all relevant users and executed.	Customer complete or enter N/A
<input type="checkbox"/> I do not have any affected devices	Customer complete or enter N/A
<input type="checkbox"/> I have a query, please contact me	Refer to contact details below

I confirm I have the following devices and confirm that the product is either new and used or currently being used with a patient (first single patient) and requires a replacement as specified in the Field Safety Notice. Please add more rows as required.

PRODUCT PS001 or PS005 or PS006	Serial Number	CURRENT LOCATION
TOTAL AFFECTED PRODUCT		

Customer Response Form

Reference: Field Safety Notice: RFA-2A

Organisation			
Contact Name / Title			
Address (no PO boxes, please)			
City, State, Zip			
Phone Number		Fax:	
E-Mail Address:			
Signature		Date	

- ☐ We have supplied potentially affected product to another organisation.
If checked : please provide new organisation information below.

New Organisation Name			
Contact Name / Title			
Address*			
City, State, Zip			
Phone Number		Fax:	
E-Mail Address:			

Address to Return Product To:

JAN KREDIET VENLO
Jacob Roggeveenweg 8-10
5928 LS Venlo
Netherlands

For the attention of: The Insides Company

Please provide it to your Distributor (if applicable) or you can request a courier bag from www.theinsides.co/returns page and we will send you a courier bag/box for returning the old unit. We will responsibly dispose of these units.

Please return your completed form by email to:

tina.mason@theinsides.co and garth.sutherland@theinsides.co