



Batch number and expiry date of the Safety Needle can be identified on the label on the box that includes 25 Safety Needles or on the label on the packaging of each Safety Needle.

Batch number of the Safety Needle on the photos above is marked with green.

Shelf life of 2 years, that needs to be replaced by the expiry date stated in the table below can be identified is marked with the red box on the photos.

Effected batch numbers of the Safety Needle are stated in the table below, including a change in the expiry date to one year.

BATCH NUMBER OF THE SAFETY NEEDLE	DATE OF MANUFACTURING	EXPIRY DATE THAT SHOULD BE TAKEN INTO ACCOUNT
20220525001	8-Jun-2022	31-May-2023
20220919001	9-Nov-2022	30-Sep-2023
20221121001	28-Nov-2022	30-Nov-2023
20230209001	16-Feb-2023	28-Feb-2024
20230321001	5-Apr-2023	30-Mar-2024

We invite you to take the following actions:

- Identify if any of the batches listed in the table are in stock
- Take into account the expiry date of one year, that is identified in the table per batch of Safety Needles

- In case needles you have in stock are expired according to the new expiry date, please notify the manufacturer in order to arrange exchange of the Safety Needles

Please acknowledge receipt of this communication by returning the Annex I as soon as possible and no later than 30-Nov-2023.

We remain at your disposal for any questions or assistance.

Manufacturer contacts:

Medical Precision B.V.

Telfordstraat 9-30

8013RL Zwolle

The Netherlands


info@medicalprecision.nl

juha@medicalprecision.nl

Zwolle, 14-Nov-2023

Roland Kortenhorst, CEO

Signature:



Annex I

FORM TO BE FILLED IN BY THE DISTRIBUTOR/ END USER

DISTRIBUTOR/END USER DETAILS	
COMPANY NAME	
ADDRESS	
CONTACT NAME	
PHONE NUMBER	
EMAIL	

PLEASE IDENTIFY THE BATCH OF SAFETY NEEDLE YOU HAVE IN STOCK AND THE QUANTITY

SAFETY NEDLE BATCH NUMBER	QUANTITY
20220525001	
20220919001	
20221121001	
20230209001	
20230321001	

PLEASE SELECT THE ACTION, YOU DECIDED TO FOLLOW:

- ☐ Batch of Safety Needles was identified decision use as is, taking into account 1-year shelf life
- ☐ Batch of Safety Needles was identified and decided for replacement of current stock – follow the instructions bellow

INSTRUCTIONS FOR REPLACEMENT OPTION

In case you decided the Safety Needles to ship back to be replaced, please ship them to the following address:

Medical Precision B.V.
Telfordstraat 9-30
8013RL Zwolle
The Netherlands

Name:

Date:

Place:

Signature:

It is important that your organization takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organization's reply is the evidence we need to monitor the progress of the corrective actions.