



THUASNE®

INFORMATION IMPORTANTE

RAPPEL DE PRODUIT

Globe-Trotter
Globe-Trotter+

Vous nous avez fait confiance en achetant nos produits Cannes Anglaises **Globe-Trotter et Globe-Trotter+** et nous vous en remercions.


Nous souhaitons vous informer d'un **risque de non-conformité** de nos dispositifs :

Cannes Anglaises **Globe-Trotter et Globe-Trotter+**

Jusqu'alors, nos produits étaient recommandés pour un poids maximum de 150 kg. **Par mesure de précaution nous avons décidé de réduire ce poids maximum à 130 kg.**

PAR CONSÉQUENT :

Si votre poids est supérieur à 130 kg, merci de ne plus utiliser nos produits et de les ramener impérativement chez votre fournisseur.



Si votre poids est inférieur à 130 kg, vous pouvez continuer d'utiliser nos produits. Cependant, merci de revenir vers votre fournisseur qui vous fournira un addendum à la notice d'utilisation ainsi que le sticker ci-contre à apposer sur le tube en aluminium.



*Nous vous prions d'accepter nos excuses pour les désagréments occasionnés par cette opération. Plus que jamais nous avons à cœur d'assurer la sécurité des patients.
Nous restons à votre disposition pour toutes questions.*



©Thuisne - 0000000 - (2024-05)

Thuisne
120, rue Marius Aufan
92300 Levallois-Perret (France)
www.thuisne.fr



Réserve pour labels certif. papier



THUASNE®

IMPORTANT INFORMATION PRODUCT RECALL

Globe-Trotter Globe-Trotter+


You trusted us by choosing our **Globe-Trotter and Globe-Trotter+** products and we thank you.

We would like to inform you of a **risk of non-compliance** with our **Globe-Trotter and Globe-Trotter+** crutches

Until now, our products were recommended for a maximum weight of 150 kg. **As a precautionary measure we have decided to reduce this maximum weight to 130 kg.**

CONSEQUENTLY:

If your weight is greater than 130 kg, please no longer use our products and bring them back to your supplier.



If your weight is less than 130 kg, you can continue to use our products. However, please contact your supplier who will provide you with an addendum to the instructions for use as well as the sticker to affix to the aluminium tube.

W2016021010
Globe-Trotter



W2016022010
Globe-Trotter



W2017022006
Globe-Trotter+



Please accept our apologies for the inconvenience caused by this operation. More than ever, we are committed to ensuring patient safety.
We remain at your disposal for any questions.



Thuasne
120, rue Marius Aufan
92300 Levallois-Perret (France)
www.thuasne.com

©Thuasne - 0000000 - (2024-05)

